

OPERATING INSTRUCTIONS AND SYSTEM DESCRIPTION OF THE

VA-10

VOLTAMMETRIC AND AMPEROMETRIC AMPLIFIER FOR EPMS SYSTEMS

with 4-pole BESSEL Filter



VERSION 3.2

npi 2007

Table of Contents

1. Safety Regulations.....	3
2. EPMS-07 Modular Plug-In System.....	4
2.1. Components of the EPMS-07 Housing	4
2.2. General System Description / Operation	4
2.3. System Grounding	5
2.4. Technical Data.....	5
3. Introduction	6
4. VA-10 Components.....	7
5. VA-10 System	7
5.1. System Description.....	7
5.2. Description of the Front Panel.....	8
6. Headstage	11
6.1. Headstage Elements.....	11
7. Operation	12
7.1. Setting up the VA-10.....	12
7.2. Carbon-Fiber Electrodes.....	13
7.3. Reference- / Counterelectrode.....	14
7.4. Amperometric Measurements.....	14
7.5. Cyclic Voltammetry	14
8. Literature	15
VA-10 used for lipid bilayers	17
9. Technical Data.....	19

1. Safety Regulations

VERY IMPORTANT: Instruments and components supplied by npi electronic are **NOT** intended for clinical use or medical purposes (e.g. for diagnosis or treatment of humans), or for any other life-supporting system. npi electronic disclaims any warranties for such purpose. Equipment supplied by npi electronic must be operated only by selected, trained and adequately instructed personnel. For details please consult the **GENERAL TERMS OF DELIVERY AND CONDITIONS OF BUSINESS** of npi electronic, D-71732 Tamm, Germany.

- 1) **GENERAL:** This system is designed for use in scientific laboratories and must be operated by trained staff only. General safety regulations for operating electrical devices are to be followed.
- 2) **AC MAINS CONNECTION:** While working with the npi systems, always adhere to the appropriate safety measures for handling electronic devices. Before using any device please read manuals and instructions carefully.
The device is to be operated only at 115/230 Volt 60/50 Hz AC. Please check for appropriate line voltage before connecting any system to mains.
Always use a three-wire line cord and a mains power-plug with a protection contact connected to ground (protective earth).
Before opening the cabinet unplug the instrument.
Unplug the instrument when replacing the fuse or changing line voltage. Replace fuse only with an appropriate specified type.
- 3) **STATIC ELECTRICITY:** Electronic equipment is sensitive to static discharges. Some devices such as sensor inputs are equipped with very sensitive FET amplifiers, which can be damaged by electrostatic charge and must therefore be handled with care. Electrostatic discharge can be avoided by touching a grounded metal surface when changing or adjusting sensors. **Always turn power off when adding or removing modules, connecting or disconnecting sensors, headstages or other components from the instrument or 19" cabinet.**
- 4) **TEMPERATURE DRIFT / WARM-UP TIME:** All analog electronic systems are sensitive to temperature changes. Therefore, all electronic instruments containing analog circuits should be used only in a warmed-up condition (i.e. after internal temperature has reached steady-state values). In most cases a warm-up period of 20-30 minutes is sufficient.
- 5) **HANDLING:** Please protect the device from moisture, heat, radiation and corrosive chemicals.

2. EPMS-07 Modular Plug-In System

2.1. Components of the EPMS-07 Housing

The following items are shipped with the EPMS-07 housing:

- ✓ EPMS-07 cabinet
- ✓ Power cord
- ✓ Fuse 0.8 / 0.4A, slow
- ✓ Front covers

2.2. General System Description / Operation

The npI EPMS-07 is a modular system for processing of bioelectrical signals in electrophysiology (see Figure 1). The system is housed in a 19" rackmount cabinet (3U) containing a power supply and has room for up to 7 plug-in units. The plug-in units are connected to power by a bus at the rear panel.

The plug-in units must be kept in position by four screws (M 2,5 x 10). The screws are important not only for mechanical stability but also for proper electrical connection to the system housing. Free area must be protected with covers.

In order to avoid induction of electromagnetic noise the power supply unit, the power switch and the fuse are located at the rear of the housing.



Figure 1: Example of a measurement system located in an EPMS housing with two channel iontophoresis and balance module (MVCS-02, MVCC-02), bridge amplifier module (BRAMP-01R) and an extracellular amplifier module (EXT-01C)

2.3. System Grounding

The 19" cabinet is grounded by the power cable through the ground pin of the mains connector (= protective earth, see chapter 1). In order to avoid ground loops the internal ground is isolated from the protective earth. The internal ground is used on the BNC connectors or GROUND plugs of the modules that are inserted into the EPMS-07 housing. The internal ground and mains ground (= protective earth) can be connected by a wire using the ground plugs on the rear panel of the instrument. It is not possible to predict whether measurements will be less or more noisy with the internal ground and mains ground connected. We recommend that you try both arrangements to determine the best configuration.

2.4. Technical Data

- 19" rackmount cabinet, 3U high (1U=1 3/4" = 44.45 mm), for up to 7 plug-in units
- Power supply: 115/230 Volts AC, 60/50 Hz, fuse 2 A / 1 A slow, 45-60 W

3. Introduction

Recently, electrochemical methods using carbon-fiber microelectrodes have been applied to measure the release of oxidizable transmitter from *single cells*, and, even more impressively, from *single exocytotic vesicles*. Transmitters that are oxidizable and which, therefore, can be measured with this approach, include serotonin, dopamine, adrenaline, and noradrenaline. In addition, some peptides or proteins such as insulin may be oxidizable owing to the presence of oxidizable amino acids such as cysteine or tyrosine.

Cells that have been studied successfully with this technique include adrenal chromaffin cells (1,2), sympathetic neurons (3), mast cells (4), pancreatic beta cells (5,6,7), carotid glomus cells (8) and melanotrophs (9), but the list is growing. In addition, in brain slices simultaneous intracellular and voltammetric studies have been made to correlate intracellular electric signals with transmitter release (10).

Two useful electrochemical approaches are **amperometry** and **cyclic voltammetry**. In **amperometry**, a DC potential is applied to a carbon-fiber microelectrode. The applied potential appears at the interface between the carbon and the mammalian ringer solution. If the potential is much greater than the redox potential for a given transmitter, then molecules of transmitter diffusing to the carbon surface are oxidized rapidly yielding a current that can be measured. The sensitivity of the **amperometric** approach, in particular, has provided an unprecedented look at the time course of transmitter release revealing distinct phases of release (2, 11). On the other hand, the amperometric approach provides little information about the substance being oxidized or reduced.

Cyclic voltammetry provides a limited amount of information about the substance being studied, at some expense to the time resolution. In this approach a cyclically repeating voltage waveform, typically consisting of voltage ramps, is applied to the carbon-fiber electrode and the resulting current is plotted as a function of the applied voltage (after subtraction of a "background" record obtained in the absence of the redox species). Since different substances have different potentials for oxidation and for reduction one can distinguish transmitters from each other.

For more detailed informations about the principles of electrochemical measurements at single cells and the fabrication of carbon-fiber microelectrodes refer to several recent reviews (e.g. 12,13).

4. VA-10 Components

The following items are shipped with the VA-10 system:

- ✓ VA-10 amplifier
- ✓ Headstage
- ✓ GND connector for headstage (1 mm)
- ✓ User manual

Optional accessories:

- ⇨ Carbon-fiber electrode holder
- ⇨ Carbon-fiber electrodes (CFE), \varnothing 5 μ m
- ⇨ Headstage with differential input
- ⇨ Four VA-10 amplifier system:
Modified EPMS housing, 4 VA-10 amplifier modules, and a four-in-one headstage

5. VA-10 System

5.1. System Description

The VA-10 is a sensitive (picoampere range) current amplifier that is intended for voltammetric measurements with carbon-fiber microelectrodes in biological systems, where the total currents do not exceed 20 nA. It can be used for either DC amperometry using the built-in voltage source or it can be operated with user-supplied external voltage waveforms (e.g. for cyclic voltammetry).

The VA-10 is ideally suited for measurements with carbon-fiber disk microelectrodes having diameters of 10 μ m or less from single cells plated onto glass cover slips. However, it can also be used for measurements made on superficially located cells in tissue slices. The VA-10 is not recommended for use in *in vivo* recordings with carbon-fiber electrodes having long cylindrical measuring surfaces, because in this case currents approach the μ A range and a third electrode is required to compensate for the IR drop as currents flow through the extracellular fluid.

(2) FILT. OUTPUT connector

BNC connector providing a filtered voltage proportional to the current passed through the electrode. The signal is filtered by a low pass filter and amplified by the GAIN factor selected by switch (3). The corner frequency of the filter is selected by switch (5).

(3) GAIN mV/pA switch

6-position rotary switch to set the amplification factor (x0.5, x1, x2.5, x5, x10, x25 mV / pA) of the current proportional voltage signal at (1) and (2).

(4) OVER LEDs

LEDs that indicate if the amplifier 10% below it's positive or negative limit (± 10 V). The linear range of the amplifier is ± 12 V.

(5) LOW PASS Hz switch

6-position rotary switch to set the corner frequency of the low pass 4-pole Bessel filter. The filtered OUTPUT can be obtained at connector (2).

(6) COMMAND voltage display

LCD that indicates the COMMAND voltage applied to the electrode (range: ± 1000 mV, display: XXXX mV). The COMMAND voltage is set by potentiometer (7).

COMMAND unit

The COMMAND unit consists of (7) COMMAND potentiometer and (8) +/- switch

(7) COMMAND potentiometer

10-turn potentiometer to set the command voltage in DC amperometric experiments using the internal voltage source (range: ± 1000 mV).

(8) +/- switch

Switch to set the polarity of the COMMAND voltage. In position 0 the COMMAND voltage generated by the VA-10 is disabled.

(9) INPUT /10 connector

BNC connector to connect an external waveform for fast cyclic voltammetry. The INPUT voltage is divided by 10 internally and applied to the electrode.

Important: The voltage at the electrode is always the sum of the voltage at INPUT /10 connector and the setting at the COMMAND voltage potentiometer.

(10) COMMAND MONITOR connector

BNC connector providing the COMMAND voltage at the electrode multiplied by a factor of 10, i.e. the voltage is the sum of the setting at (7) COMMAND potentiometer times 10 and the voltage at (9) INPUT /10 connector.

(11) GAIN MONITOR connector

BNC connector that monitors the setting of (3) GAIN switch. Range: +1 to +6 V, resolution 1 V / STEP (i.e. 3V indicate a GAIN of 2).

(12) HEADSTAGE connector

8-pin connector for the HEADSTAGE cable.

Important: Always turn power off when connecting or disconnecting the headstage.

6. Headstage

The VA-10 comes with the standard headstage (range: ± 1000 mV) for connecting carbon-fiber electrodes via an electrode holder (optional).

A 3-electrode headstage with differential input (see also **Optional accessories** in chapter 4) is also available. For details contact npj.



Figure 3: Headstage of the VA-10

6.1. Headstage Elements

- 1 BNC connector for the electrode holder
- 2 GROUND: ground
- 3 COMMAND: command potential output
- 4 headstage cable to amplifier
- 5 REFERENCE: not installed

In the 2-electrode headstage the REFERENCE is not installed. The table indicates whether the headstage is equipped with the standard feedback resistor (500 M Ω) or with a different one. It is also marked whether the headstage is in 2-electrode or in 3-electrode configuration.

The electrical connections are made like in a conventional patch-clamp headstage (e.g. the headstage of the EPC-7 (Heka elektronik, Lambrecht, Germany).

The carbon-fiber electrode fits into the BNC connector of the headstage (#1, Figure 3). An electrode holder (optional) gives additional mechanical stability. Ask npj for details. GROUND provides the ground and is linked to the bath, e.g. via an Ag-AgCl pellet. COMMAND provides the command potential at the electrode and remains usually open, but

it can be used to optimize the measurements by connecting it to an electrode shield (see Ogden (1994) for setting up a driven shield configuration). The headstage is attached to the amplifier with the headstage cable (see #4, Figure 3) and an 8-pole connector. For maximal flexibility the headstage is mounted on a plastic plate by customized screws. Thus, the user can modify the mounting plate according to his needs, e.g. to mount the headstage to a micromanipulator.

Note: The shield of the BNC connector and the enclosure of the headstage are linked to the command potential output and must not be connected to ground.

Caution: Please always adhere to the appropriate safety precautions (see chapter **Fehler! Verweisquelle konnte nicht gefunden werden.**). Please turn power off when connecting or disconnecting the headstage from the HEADSTAGE connector!

7. Operation

7.1. Setting up the VA-10

The VA-10 EPMS amplifier is shipped as a plug-in unit for the EPMS-07 modular system and equipped with a small headstage with a BNC connector. When the system arrives the headstage will not be connected to the cabinet.

For biological voltammetric measurements the experimental setup typically consists of a microscope located within a Faraday cage to minimize noise pickup. A manipulator is used for positioning the voltammeter headstage with an attached electrode, so that the electrode tip is near the cell(s) to be studied.

After installing the VA-10 plug-in unit in the EPMS-07 housing (see also chapter 2.2) mount the headstage to the manipulator.

Important: Always turn power off when connecting or disconnecting the headstage.

To facilitate noise reduction of the setup, the faraday cage and the microscope may be connected to the INTERNAL GROUND located on the back of the EPMS cabinet. Needless to say, grounding for low noise is an art. If you are not familiar with the principles of low noise connections, you should consult the local electrophysiology expert or electrical engineer (see also chapter 2.3).

- ❑ Before turning on power, set the switches at the front panel to the following positions:

Gain (#3, Figure 2):	1 mV/pA
LP Filter (#5, Figure 2):	5k
+/- (#8, Figure 2):	0

- ❑ Turn power on. The reading of the displays in the modules are an indicator for a working power supply.

As mentioned above the VA-10 can be used for DC amperometry, taking advantage of the internal voltage source, or it can be used with user-supplied external waveforms, e.g. for cyclic voltammetry.

- ❑ The LCD display should read 0. Use the COMMAND potentiometer (#7, Figure 2) to apply a constant voltage for DC amperometry. The 3-position polarity switch can be set to “+” or “-”, depending on the polarity of the desired command potential. The command voltage is displayed at the LCD in mV. This voltage is applied to the electrode mounted on the headstage.
- ❑ If you intend to read the signal from the VA-10 into a data acquisition system, connect a BNC cable from the acquisition system to FILT. OUTPUT (#2, Figure 2) or UNFILTERED OUTPUT (#1, Figure 2). Additionally, you can monitor the GAIN setting by connecting a BNC cable from the acquisition system to GAIN MONITOR (#11, Figure 2).
- ❑ If you intend to use an external voltage source you need to connect your external voltage source to (9) INPUT /10 connector. Remember that the input voltage will be scaled down by a factor of 10 at the headstage. Note that, when an external voltage source is used, the 3-position toggle switch controlling the internal voltage source should be set to “0”, unless you want to sum the external voltage with the internal voltage source.

Important: The voltage at the electrode is always the sum of the voltage at INPUT /10 connector and the setting at the COMMAND voltage potentiometer.

The VA-10 is now ready for measurements.

7.2. Carbon-Fiber Electrodes

Most voltammetric measurements in today’s biological investigations involve the use of carbon-fiber electrodes. These electrodes can be purchased or you can make your own. For use with the VA-10 voltammeter the electrodes must fit to the BNC connector at the input of the amplifier. Two types of connection are commonly used:

- 1) direct connection via a BNC pin that is soldered onto the end of the electrode or
- 2) connection via a metal/liquid junction, for example using a 3 M KCl solution to interface the end of a carbon fiber to a Ag/AgCl wire.

For the first type of connection no special holder is required. For the metal/liquid junction type a special electrode holder must be used. For some electrodes a patch-pipette holder is adequate.

Carbon fiber disk microelectrodes with small diameter (5-10 μm range) can be obtained from npi or ALA Scientific Instruments. The electrodes are manufactured using an anodic electrophoretic insulation method (Schulte, A. and R. Chow, 1996, Anal. Chem. 68, 3054-3058).

7.3. Reference- / Counterelectrode

The counter electrode used for biological measurements is typically a Ag/AgCl pellet (a sodium-saturated calomel electrode is sometimes used). The pellet should be immersed into the recording chamber and connected via a thin wire to the ground input of the headstage (#2, Figure 3).

7.4. Amperometric Measurements

For high time resolution measurements of transmitter release from single vesicles DC amperometry is the appropriate approach. In this approach, the carbon-fiber electrode is energized with a command potential that exceeds the redox potential of the transmitter being studied. In practice, a command potential of equal to or greater than +650 mV is sufficient for measurements of all major oxidizable transmitters that have been studied to date.

When generating a command potential for DC amperometry, there should be no control voltage at the INPUT /10 BNC. The 3-position command toggle switch (#8, Figure 2) should be set, for example to the “+” position. Then, the desired potential can be dialed in with the 10-turn potentiometer. As indicated above +650 mV is sufficient for most measurements with cells.

The amperometric signal is diffusion based. Thus, the distance between the carbon-fiber electrode detecting face and the cell surface must be kept to a minimum. For maximum signal size and most rapid kinetics it is possible to touch the cell membrane with the electrode.

7.5. Cyclic Voltammetry

In order to facilitate the identification of the transmitter being released, it is possible to use various voltage waveforms. One common approach is to apply fast voltage ramp potentials, i.e. to perform fast cyclic voltammetry.

For this application, it is necessary to use an external voltage source connected to the INPUT /10 connector (#9, Figure 2) at the front panel of the VA-10. Because one has to relate the measured current to the applied instantaneous voltage, the current and the applied voltage should be recorded simultaneously with a data acquisition system.

8. Literature

VA-10 typical recordings

- ❑ Bai, J., Wang, C. T., Richards, D. A., Jackson, M. B., & Chapman, E. R. (2004). Fusion pore dynamics are regulated by synaptotagmin**t*-SNARE interactions. *Neuron* **41**, 929-942.
- ❑ Barclay, J. W., Craig, T. J., Fisher, R. J., Ciufu, L. F., Evans, G. J., Morgan, A., & Burgoyne, R. D. (2003). Phosphorylation of Munc18 by protein kinase C regulates the kinetics of exocytosis. *J Biol.Chem.* **278**, 10538-10545.
- ❑ Bertrand, P. P. (2006). Real-time measurement of serotonin release and motility in guinea pig ileum. *J Physiol.* **577**, 689-704.
- ❑ Bristol, A. S., Sutton, M. A., & Carew, T. J. (2004). Neural circuit of tail-elicited siphon withdrawal in aplysia. I. Differential lateralization of sensitization and dishabituation. *Journal of Neurophysiology* **91**, 666-677.
- ❑ Bristol, A. S., Marinesco, S., & Carew, T. J. (2004). Neural Circuit of Tail-Elicited Siphon Withdrawal in Aplysia. II. Role of Gated Inhibition in Differential Lateralization of Sensitization and Dishabituation. *Journal of Neurophysiology* **91**, 678-692.
- ❑ Britt, J. P. & McGehee, D. S. (2008). Presynaptic opioid and nicotinic receptor modulation of dopamine overflow in the nucleus accumbens. *Journal of Neuroscience* **28**, 1672-1681.
- ❑ Chan, S. A., Chow, R., & Smith, C. (2003). Calcium dependence of action potential-induced endocytosis in chromaffin cells. *Pflugers Arch.* **445**, 540-546.
- ❑ Chan, S. A., Polo-Parada, L., Landmesser, L. T., & Smith, C. (2005). Adrenal chromaffin cells exhibit impaired granule trafficking in NCAM knockout mice. *J Neurophysiol.* **94**, 1037-1047.
- ❑ Ciufu, L. F., Barclay, J. W., Burgoyne, R. D., & Morgan, A. (2005). Munc18-1 Regulates Early and Late Stages of Exocytosis via Syntaxin-independent Protein Interactions. *Molecular Biology of the Cell* **16**, 470-482.
- ❑ Doreian, B. W., Fulop, T. G., & Smith, C. B. (2008). Myosin II activation and actin reorganization regulate the mode of quantal exocytosis in mouse adrenal chromaffin cells. *Journal of Neuroscience* **28**, 4470-4478.
- ❑ Evans, G. J., Barclay, J. W., Prescott, G. R., Jo, S. R., Burgoyne, R. D., Birnbaum, M. J., & Morgan, A. (2006). Protein kinase B/Akt is a novel cysteine string protein kinase that regulates exocytosis release kinetics and quantal size. *J Biol.Chem.* **281**, 1564-1572.
- ❑ Fischer, R. J., Pevsner, J., & Burgoyne, R. D. (2001). Control of Fusion Pore Dynamics During Exocytosis by Munc18. *Science* **291**, 875-878.
- ❑ Fulop, T., Radabaugh, S., & Smith, C. (2005). Activity-dependent differential transmitter release in mouse adrenal chromaffin cells. *J Neurosci.* **25**, 7324-7332.
- ❑ Graham, M. E., & Burgoyne, R. D. (2000). Comparison of Cysteine String Protein (Csp) and Mutant α -SNAP Overexpression Reveals a Role for Csp in Late Steps of Membrane Fusion in Dense-Core Granule Exocytosis in Adrenal Chromaffin Cells. *J.Neurosci.* **20**, 1281-1289.
- ❑ Graham, M. E., Barclay, J. W., & Burgoyne, R. D. (2004). Syntaxin/Munc18 interactions in the late events during vesicle fusion and release in exocytosis. *Journal of Biological Chemistry* M400827200.

- ❑ Han, X., Wang, C. T., Bai, J., Chapman, E. R., & Jackson, M. B. (2004). Transmembrane segments of syntaxin line the fusion pore of Ca²⁺-triggered exocytosis. *Science* **304**, 289-292.
- ❑ Han, X. & Jackson, M. B. (2005). Electrostatic Interactions between the Syntaxin Membrane Anchor and Neurotransmitter Passing through the Fusion Pore. *Biophys.J.* **88**, L20-L22.
- ❑ Jaffe, E. H, Marty, A., Schulte, A. and Chow, R.H. (1998). Extrasynaptic vesicular transmitter release from the somata of substantia nigra neurons in rat midbrain slices. *J.Neurosci.* **18**, 3548-3553.
- ❑ Lerner, I., Trus, M., Cohen, R., Yizhar, O., Nussinovitch, I., & Atlas, D. (2006). Ion interaction at the pore of Lc-type Ca²⁺ channel is sufficient to mediate depolarization-induced exocytosis. *J Neurochem.* **97**, 116-127.
- ❑ Moore, J. M., Papke, J. B., Cahill, A. L., & Harkins, A. B. (2006). Stable gene silencing of synaptotagmin I in rat PC12 cells inhibits Ca²⁺-evoked release of catecholamine. *Am.J Physiol Cell Physiol.* **291**, C270-C281.
- ❑ Petzinger, G. M., Walsh, J. P., Akopian, G., Hogg, E., Abernathy, A., Arevalo, P., Turnquist, P., Vuckovic, M., Fisher, B. E., Togasaki, D. M., & Jakowec, M. W. (2007). Effects of treadmill exercise on dopaminergic transmission in the 1-methyl-4-phenyl-1,2,3,6-tetrahydropyridine-lesioned mouse model of basal ganglia injury. *J Neurosci.* **27**, 5291-5300.
- ❑ Souvannakitti, D., Kumar, G. K., Fox, A., & Prabhakar, N. R. (2009). Neonatal Intermittent Hypoxia Leads to Long-Lasting Facilitation of Acute Hypoxia-evoked Catecholamine Secretion from Rat Chromaffin Cells. *J.Neurophysiol.*
- ❑ Wang, P., Wang, C. T., Bai, J., Jackson, M. B., & Chapman, E. R. (2003). Mutations in the effector binding loops in the C2A and C2B domains of synaptotagmin I disrupt exocytosis in a nonadditive manner. *J Biol.Chem.* **278**, 47030-47037.
- ❑ Wang, C. T., Bai, J., Chang, P. Y., Chapman, E. R., & Jackson, M. B. (2006). Synaptotagmin-Ca²⁺ triggers two sequential steps in regulated exocytosis in rat PC12 cells: fusion pore opening and fusion pore dilation. *J Physiol.* **570**, 295-307.
- ❑ Wang, H., Chan, S. A., Ogier, M., Hellard, D., Wang, Q., Smith, C., & Katz, D. M. (2006). Dysregulation of brain-derived neurotrophic factor expression and neurosecretory function in Mecp2 null mice. *J Neurosci.* **26**, 10911-10915.
- ❑ Xie, Z., Herring, B. E., & Fox, A. P. (2006). Excitatory and Inhibitory Actions of Isoflurane in Bovine Chromaffin Cells. *Journal of Neurophysiology*
- ❑ Xu, J., Xu, Y., Ellis-Davies, G. C. R., Augustine, G. J. & Tse, F. W. (2002). Differential Regulation of Exocytosis by α - and β -SNAPs. *J.Neurosci.* **22**, 53–61.
- ❑ Zhuge, R., Decrescenzo, V., Sorrentino, V., Lai, F. A., Tuft, R. A., Lifshitz, L. M., Lemos, J. R., Smith, C., Fogarty, K. E., & Walsh, J. V., Jr. (2006). Syntillas release Ca²⁺ at a site different from the microdomain where exocytosis occurs in mouse chromaffin cells. *Biophys.J.* **90**, 2027-2037.

VA-10 used for recordings with 3 electrodes

- ❑ Marinesco, S., & Carew, T. J. (2002). Serotonin Release Evoked by Tail Nerve Stimulation in the CNS of *Aplysia*: Characterization and Relationship to Heterosynaptic Plasticity. *J.Neurosci.* **22**, 2299–2312.

- ❑ Marinesco, S., & Carew, T. J. (2002). Improved electrochemical detection of biogenic amines in *Aplysia* using base-hydrolyzed cellulose-coated carbon fiber microelectrodes. *J.Neurosci.Meth.* **117**, 87-97.
- ❑ Marinesco, S., Kolkman, K. E., & Carew, T. J. (2004). Serotonergic modulation in *aplysia*. I. Distributed serotonergic network persistently activated by sensitizing stimuli. *J Neurophysiol.* **92**, 2468-2486.
- ❑ Marinesco, S., Wickremasinghe, N., Kolkman, K. E., & Carew, T. J. (2004). Serotonergic modulation in *aplysia*. II. Cellular and behavioral consequences of increased serotonergic tone. *J Neurophysiol.* **92**, 2487-2496.
- ❑ Marinesco, S., Wickremasinghe, N., & Carew, T. J. (2006). Regulation of behavioral and synaptic plasticity by serotonin release within local modulatory fields in the CNS of *Aplysia*. *J Neurosci.* **26**, 12682-12693.

VA-10 used for recordings with electrode arrays

- ❑ Dias, A. F., Dernick, G., Valero, V., Yong, M. G., James, C. D., Craighead, H. G., & Lindau, M. (2002). An electrochemical detector array to study cell biology on the nanoscale. *Nanotechnology* **13**, 285-289.
- ❑ Hafez, I., Kisler, K., Berberian, K., Dernick, G., Valero, V., Yong, M. G., Craighead, H. G., & Lindau, M. (2005). Electrochemical imaging of fusion pore openings by electrochemical detector arrays. *Proc.Natl.Acad.Sci.U.S.A* **102**, 13879-13884.
- ❑ Spégel, C., Heiskanen, A., Acklid, J., Wolff, A., Taboryski, R., Emnéus, J., & Ruzgas, T. (2007). On-Chip Determination of Dopamine Exocytosis Using Mercaptopropionic Acid Modified Microelectrodes. *Electroanalysis* **19**, 263-271.

VA-10 used for scanning electrochemical microscopy

- ❑ Etienne, M., Schulte, A., & Schuhmann, W. (2004). High resolution constant-distance mode alternating current scanning electrochemical microscopy (AC-SECM). *Electrochem.Comm.* **6**, 288-293.
- ❑ Hengstenberg, A., Dietzel, I. D., & Schuhmann, W. (1999). Visualization of biological activities using the scanning electrochemical microscope. In: *Monitoring Molecules in Neuroscience*. ed.: Rollema, H., Abercombie, E., Sulzer, D., & Zackheim, J., Proceedings of the 8th international conference on *in vivo* methods, 19-23 June 1999, S.U.N.Y. at Stony Brook, New York, USA, 47-48.
- ❑ Hengstenberg, A., Dietzel, I. D., Blöchl, A., & Schuhmann, W. (1999). Zell-Zell-Kommunikationsprozesse mittels elektrochemischer Rastermikroskopie. *BioForum* **10**, 595-599, GIT Verlag, Darmstadt, Germany.
- ❑ Turcu, F., Schulte, A., Hartwich, G., & Schuhmann, W. (2004). Label-Free Electrochemical Recognition of DNA Hybridization by Means of Modulation of the Feedback Current in SECM. *Angew.Chem.Int.Ed Engl.* **43**, 3482-3485.

VA-10 used for lipid bilayers

- ❑ Horner, A., Antonenko, Y. N., & Pohl, P. (2009). Coupled diffusion of peripherally bound peptides along the outer and inner membrane leaflets. *Biophys.J.* **96**, 2689-2695.

References (methods)

- ❑ Alvarez de Toledo, G., Fernandez-Chacon, R., & Fernandez, J. M. (1993). Release of secretory products during transient vesicle fusion. *Nature* **363**, 554-557.
- ❑ Chow, R. H., von Rüden, L., & Neher, E. (1992). Delay in vesicle fusion revealed by electrochemical monitoring of single secretory events in adrenal chromaffin cells. *Nature* **356**, 60-63.
- ❑ Chow, R. H., & von Rüden, L. (1995). Electrochemical Detection of Secretion from Single Cells, in: Sakmann, B., & Neher, E. (eds.). *Single Channel Recording*. Plenum Press, New York and London.
- ❑ Huang, L., Shen, H., Atkinson, M. A., & Kennedy, R. T. (1995). Detection of exocytosis at individual pancreatic beta-cells by amperometry at a chemically-modified microelectrode. *Proc.Natl.Acad.Sci.* **92**, 9608-9612.
- ❑ Kudernatsch, M., & Sutor, B. (1994). Cholinergic modulation of dopamine overflow in the rat neo-striatum: a fast cyclic voltammetric study in vitro. *Neurosci.Lett.* **181**, 107-112.
- ❑ Paras, C. D., & Kennedy, R. T., (1995). Electrochemical detection of exocytosis at single-rat melanotrophs. *Anal.Chem.* **67**, 3633-3637.
- ❑ Smith, P. A., Duchon, M. R., & Ashcroft, F. M., (1995). Fluorometric and amperometric study of calcium and secretion in isolated mouse pancreatic beta-cells. *PflugersArch.* **430**, 808-818.
- ❑ Taylor, A. R., & Chow, H. (2001). A microelectrochemical technique to measure trans-plasma membrane electron transport in plant tissue and cells *in vivo*, *PlantCellEnviron*, **24**, 1-6.
- ❑ Urena, J., Fernandez-Chacon, R., Benot, A. R., Alvarez de Toledo, G., & Lopez-Barneo, J., (1994). Hypoxia induces voltage-dependent Ca²⁺ entry and quantal dopamine secretion in carotid body glomus cells. *Proc.Natl.Acad.Sci. USA* **91**, 10208-10211.
- ❑ Wightman, R.M., J.A. Jankowski, R.T. Kennedy, K.T. Kawagoe, T.J. Schroeder, D.J Leszczyszyn, J.A. Near, E.J. Dilberto, Jr., and O.H. Viveros. 1991. Temporally resolved catecholamine spikes correspond to single vesicle release from individual chromaffin cells. *Proc.Natl.Acad.Sci. USA* **88**, 10754-10758.
- ❑ Zhou, Z, Mislser, S., & Chow, R. H. (1996). Rapid fluctuations in transmitter release from single vesicles in bovine adrenal chromaffin cells. *Biophys.J.* **70**, 1543-1552.
- ❑ Zhou, Z., & Mislser, S., (1995). Amperometric detection of stimulus-induced quantal release of catecholamines from cultured superior cervical-ganglion neurons. *Proc.Natl.Acad.Sci. USA* **92**, 6938-6942.
- ❑ Zhou, Z., & Mislser, S. (1996). Amperometric detection of quantal secretion from patch-clamped rat pancreatic beta-cells. *J.Biol.Chem.* **271**, 270-277.

9. Technical Data

Headstage:

Input voltage range:	±1200 mV
Operating voltage:	±12 V
Enclosure:	Size: 40 x 70 x 20 mm, grounded mounting plate: 50 x 70 mm, not conducting
Electrode connector:	BNC with driven shield
Ground connector:	1 mm connector (black)
Driven shield output:	1 mm connector (red)
Input resistance:	>10 ¹³ Ω
Current range:	±20 nA max.
Feedback resistor:	500 MΩ

Noise: <1 pA

LP FILTER: 4-pole Bessel FILTER, attenuation: -24 dB/octave, corner frequencies: 100, 300, 500, 1k, 3k and 5k Hz

GAIN: 6-position rotary switch (x0.5, x1, x2.5, x5, x10, x25 mV / pA)

Voltage source: COMMAND VOLTAGE set by 10-turn potentiometer, range: ±1000 mV, polarity selectable with polarity switch (+, 0, -), display XXXX mV

OUTPUTS:

Resistance:	250 Ω
FILTERED:	BNC connector, sensitivity selectable with GAIN switch, FILTERED with LP Bessel FILTER
UNFILTERED:	BNC connector, sensitivity selectable with GAIN switch

INPUTS:

Resistance:	1 MΩ
COMMAND VOLTAGE:	INPUT BNC connector, sensitivity: /10

MONITORING:

Impedance:	250 Ω
GAIN:	BNC connector, 1 V per step, +1 V...+6 V
COMMAND:	BNC connector, sensitivity: x10

EPMS-07 SYSTEM

Power requirements: 115/230 V AC, 60/50 Hz, fuse 2 A / 1 A, slow, 45-60 W (dependent on the modules plugged in)

Dimensions: 19" rackmount cabinet, 3U high (1U = 1 3/4" = 44.45 mm)